



Ministry of Health and Long-Term Care

Inspection Report under the Long-Term Care Homes Act, 2007

Ministère de la Santé et des Soins de longue durée

Rapport d'inspection prévue le Loi de 2007 les foyers de soins de longue

Health System Accountability and Performance
Division
Performance Improvement and Compliance Branch
Division de la responsabilisation et de la performance du système de santé
Direction de l'amélioration de la performance et de la conformité

Ottawa Service Area Office
347 Preston St, 4th Floor
OTTAWA, ON, K1S-3J4
Telephone: (613) 569-5602
Facsimile: (613) 569-9670

Bureau régional de services d'Ottawa
347, rue Preston, 4^{ème} étage
OTTAWA, ON, K1S-3J4
Téléphone: (613) 569-5602
Télécopieur: (613) 569-9670

Public Copy/Copie du public

Date(s) of inspection/Date(s) de l'inspection	Inspection No/ No de l'inspection	Type of Inspection/Genre d'inspection
Jul 25, Aug 3, 2011	2011_036126_0017	Complaint

Licensee/Titulaire de permis

1663432 ONTARIO LTD.
2212 GLADWIN CRESCENT, UNIT A-9, SUITE 200, OTTAWA, ON, K1B-5N1

Long-Term Care Home/Foyer de soins de longue durée

MANOIR MAROCHEL
949 MONTREAL ROAD, OTTAWA, ON, K1K-0S6

Name of Inspector(s)/Nom de l'inspecteur ou des inspecteurs

LINDA HARKINS (126)

Inspection Summary/Résumé de l'inspection

The purpose of this inspection was to conduct a Complaint inspection.

During the course of the inspection, the inspector(s) spoke with the Administrator, the Director of care, the Clinical Nurse and two Health Care Aides.

During the course of the inspection, the inspector(s) Reviewed the resident health record, reviewed the transfer method with staff and ensure the lift used at the time of the incident was functioning properly.

The following Inspection Protocols were used in part or in whole during this inspection:

Accommodation Services - Maintenance

Personal Support Services

Findings of Non-Compliance were found during this inspection.

NON-COMPLIANCE / NON-RESPECT DES EXIGENCES

Definitions

WN – Written Notification
VPC – Voluntary Plan of Correction
DR – Director Referral
CO – Compliance Order
WAO – Work and Activity Order

Définitions

WN – Avis écrit
VPC – Plan de redressement volontaire
DR – Aiguillage au directeur
CO – Ordre de conformité
WAO – Ordres : travaux et activités

Non-compliance with requirements under the Long-Term Care Homes Act, 2007 (LTCHA) was found. (A requirement under the LTCHA includes the requirements contained in the items listed in the definition of "requirement under this Act" in subsection 2(1) of the LTCHA.)

The following constitutes written notification of non-compliance under paragraph 1 of section 152 of the LTCHA.

Le non-respect des exigences de la Loi de 2007 sur les foyers de soins de longue durée (LFSLD) a été constaté. (Une exigence de la loi comprend les exigences qui font partie des éléments énumérés dans la définition de « exigence prévue par la présente loi », au paragraphe 2(1) de la LFSLD.

Ce qui suit constitue un avis écrit de non-respect aux termes du paragraphe 1 de l'article 152 de la LFSLD.

WN #1: The Licensee has failed to comply with O.Reg 79/10, s. 90. Maintenance services

Specifically failed to comply with the following subsections:

- s. 90. (2) The licensee shall ensure that procedures are developed and implemented to ensure that,**
- (a) electrical and non-electrical equipment, including mechanical lifts, are kept in good repair, and maintained and cleaned at a level that meets manufacturer specifications, at a minimum;**
 - (b) all equipment, devices, assistive aids and positioning aids in the home are kept in good repair, excluding the residents' personal aids or equipment;**
 - (c) heating, ventilation and air conditioning systems are cleaned and in good state of repair and inspected at least every six months by a certified individual, and that documentation is kept of the inspection;**
 - (d) all plumbing fixtures, toilets, sinks, grab bars and washroom fixtures and accessories are maintained and kept free of corrosion and cracks;**
 - (e) gas or electric fireplaces and heat generating equipment other than the heating system referred to in clause (c) are inspected by a qualified individual at least annually, and that documentation is kept of the inspection;**
 - (f) hot water boilers and hot water holding tanks are serviced at least annually, and that documentation is kept of the service;**
 - (g) the temperature of the water serving all bathtubs, showers, and hand basins used by residents does not exceed 49 degrees Celsius, and is controlled by a device, inaccessible to residents, that regulates the temperature;**
 - (h) immediate action is taken to reduce the water temperature in the event that it exceeds 49 degrees Celsius;**
 - (i) the temperature of the hot water serving all bathtubs and showers used by residents is maintained at a temperature of at least 40 degrees Celsius;**
 - (j) if the home is using a computerized system to monitor the water temperature, the system is checked daily to ensure that it is in good working order; and**
 - (k) if the home is not using a computerized system to monitor the water temperature, the water temperature is monitored once per shift in random locations where residents have access to hot water. O. Reg. 79/10, s. 90 (2).**

Findings/Faits sayants :

1. The Home does not have any policy and procedure related to methods to reduce risk and monitor outcomes for malfunctioning equipment to ensure equipment is kept in good repair.
2. The Administrator stated that they had a few incidents with their lifts malfunctioning and was unable to confirm which lift was/were involve in the incident(s).[s.90.(2)(a)]

Issued on this 3rd day of August, 2011



**Ministry of Health and
Long-Term Care**

**Inspection Report under
the Long-Term Care
Homes Act, 2007**

**Ministère de la Santé et des
Soins de longue durée**

**Rapport d'inspection
prévus le Loi de 2007 les
foyers de soins de longue**

Signature of Inspector(s)/Signature de l'inspecteur ou des inspecteurs