



**Ministry of Health and
Long-Term Care**

**Inspection Report under
the Long-Term Care
Homes Act, 2007**

**Ministère de la Santé et des
Soins de longue durée**

**Rapport d'inspection sous la
Loi de 2007 sur les foyers de
soins de longue durée**

**Health System Accountability and
Performance Division
Performance Improvement and
Compliance Branch**

**Division de la responsabilisation et de la
performance du système de santé
Direction de l'amélioration de la
performance et de la conformité**

London Service Area Office
130 Dufferin Avenue 4th floor
LONDON ON N6A 5R2
Telephone: (519) 873-1200
Facsimile: (519) 873-1300

Bureau régional de services de
London
130 avenue Dufferin 4ème étage
LONDON ON N6A 5R2
Téléphone: (519) 873-1200
Télécopieur: (519) 873-1300

Public Copy/Copie du public

Report Date(s) / Date(s) du apport	Inspection No / No de l'inspection	Log # / Registre no	Type of Inspection / Genre d'inspection
Aug 12, 2015	2015_418615_0019	008269-15	Follow up

Licensee/Titulaire de permis

MEADOW PARK (LONDON) INC
689 YONGE STREET MIDLAND ON L4R 2E1

Long-Term Care Home/Foyer de soins de longue durée

MEADOW PARK (LONDON) INC.
1210 SOUTHDALE ROAD EAST LONDON ON N6E 1B4

Name of Inspector(s)/Nom de l'inspecteur ou des inspecteurs

HELENE DESABRAIS (615)

Inspection Summary/Résumé de l'inspection



**Ministry of Health and
Long-Term Care**

**Ministère de la Santé et des
Soins de longue durée**

**Inspection Report under
the Long-Term Care
Homes Act, 2007**

**Rapport d'inspection sous la
Loi de 2007 sur les foyers de
soins de longue durée**

The purpose of this inspection was to conduct a Follow up inspection.

This inspection was conducted on the following date(s): July 28 and 29, 2015

This follow up inspection was completed related to inspection #2015_418615_003 where a Compliance Order was issued under s.153. (1)(a).

During the course of the inspection, the inspector(s) spoke with the Administrator, the Environmental Service Supervisor and the Restorative Care Coordinator.

The inspector also reviewed bed assessment documents and other relevant documentation.

**The following Inspection Protocols were used during this inspection:
Safe and Secure Home**

During the course of this inspection, Non-Compliances were issued.

1 WN(s)

0 VPC(s)

1 CO(s)

0 DR(s)

0 WAO(s)



NON-COMPLIANCE / NON - RESPECT DES EXIGENCES

<p>Legend</p> <p>WN – Written Notification VPC – Voluntary Plan of Correction DR – Director Referral CO – Compliance Order WAO – Work and Activity Order</p>	<p>Legendé</p> <p>WN – Avis écrit VPC – Plan de redressement volontaire DR – Aiguillage au directeur CO – Ordre de conformité WAO – Ordres : travaux et activités</p>
<p>Non-compliance with requirements under the Long-Term Care Homes Act, 2007 (LTCHA) was found. (a requirement under the LTCHA includes the requirements contained in the items listed in the definition of "requirement under this Act" in subsection 2(1) of the LTCHA).</p> <p>The following constitutes written notification of non-compliance under paragraph 1 of section 152 of the LTCHA.</p>	<p>Le non-respect des exigences de la Loi de 2007 sur les foyers de soins de longue durée (LFSLD) a été constaté. (une exigence de la loi comprend les exigences qui font partie des éléments énumérés dans la définition de « exigence prévue par la présente loi », au paragraphe 2(1) de la LFSLD.</p> <p>Ce qui suit constitue un avis écrit de non-respect aux termes du paragraphe 1 de l'article 152 de la LFSLD.</p>

**WN #1: The Licensee has failed to comply with O.Reg 79/10, s. 15. Bed rails
Specifically failed to comply with the following:**

- s. 15. (1) Every licensee of a long-term care home shall ensure that where bed rails are used,**
- (a) the resident is assessed and his or her bed system is evaluated in accordance with evidence-based practices and, if there are none, in accordance with prevailing practices, to minimize risk to the resident; O. Reg. 79/10, s. 15 (1).**
 - (b) steps are taken to prevent resident entrapment, taking into consideration all potential zones of entrapment; and O. Reg. 79/10, s. 15 (1).**
 - (c) other safety issues related to the use of bed rails are addressed, including height and latch reliability. O. Reg. 79/10, s. 15 (1).**

Findings/Faits saillants :

1. The licensee has failed to ensure that where bed rails were used, the resident was assessed and his or her bed system evaluated in accordance with evidence-based practices, and if there were none, in accordance with prevailing practices to minimize risk to the resident; and to ensure that where bed rails were used, steps were taken to prevent resident entrapment, taking into consideration all potential zones of entrapment.

Review of documentation of the bed system entrapment assessments revealed assessments had not been completed for all bed systems in the home. There was no documented evidence that residents, who used bed rails, had been assessed.

Interview with the Environmental Service Supervisor (ESS) revealed that all the beds had not been assessed.

Interview with the Restorative Care Coordinator and the Administrator confirmed that the bed system entrapment assessments were not totally completed, that residents who used bed rails had not been assessed and no short term action was taken to prevent resident entrapment. [s. 15. (1)]

Additional Required Actions:

CO # - 001 will be served on the licensee. Refer to the "Order(s) of the Inspector".



**Ministry of Health and
Long-Term Care**

**Ministère de la Santé et des
Soins de longue durée**

**Inspection Report under
the Long-Term Care
Homes Act, 2007**

**Rapport d'inspection sous la
Loi de 2007 sur les foyers de
soins de longue durée**

Issued on this 13th day of August, 2015

Signature of Inspector(s)/Signature de l'inspecteur ou des inspecteurs

Original report signed by the inspector.



**Ministry of Health and
Long-Term Care**

**Ministère de la Santé et
des Soins de longue durée**

Order(s) of the Inspector

Pursuant to section 153 and/or
section 154 of the *Long-Term Care
Homes Act, 2007*, S.O. 2007, c.8

Ordre(s) de l'inspecteur

Aux termes de l'article 153 et/ou
de l'article 154 de la *Loi de 2007 sur les foyers
de soins de longue durée*, L.O. 2007, chap. 8

**Health System Accountability and Performance Division
Performance Improvement and Compliance Branch**

**Division de la responsabilisation et de la performance du système de santé
Direction de l'amélioration de la performance et de la conformité**

Public Copy/Copie du public

Name of Inspector (ID #) /

Nom de l'inspecteur (No) : HELENE DESABRAIS (615)

Inspection No. /

No de l'inspection : 2015_418615_0019

Log No. /

Registre no: 008269-15

Type of Inspection /

Genre

Follow up

d'inspection:

Report Date(s) /

Date(s) du Rapport : Aug 12, 2015

Licensee /

Titulaire de permis : MEADOW PARK (LONDON) INC
689 YONGE STREET, MIDLAND, ON, L4R-2E1

LTC Home /

Foyer de SLD : MEADOW PARK (LONDON) INC.
1210 SOUTHDALE ROAD EAST, LONDON, ON,
N6E-1B4

Name of Administrator /

Nom de l'administratrice

ou de l'administrateur : Melanie Smith

To MEADOW PARK (LONDON) INC, you are hereby required to comply with the following order(s) by the date(s) set out below:



Order(s) of the Inspector

Pursuant to section 153 and/or
section 154 of the *Long-Term Care
Homes Act, 2007, S.O. 2007, c.8*

Ordre(s) de l'inspecteur

Aux termes de l'article 153 et/ou
de l'article 154 de la *Loi de 2007 sur les foyers
de soins de longue durée, L.O. 2007, chap. 8*

Order # /
Ordre no : 001 **Order Type /**
Genre d'ordre : Compliance Orders, s. 153. (1) (a)

Linked to Existing Order /
Lien vers ordre 2015_418615_0003, CO #001;
existant:

Pursuant to / Aux termes de :

O.Reg 79/10, s. 15. (1) Every licensee of a long-term care home shall ensure that where bed rails are used,
(a) the resident is assessed and his or her bed system is evaluated in accordance with evidence-based practices and, if there are none, in accordance with prevailing practices, to minimize risk to the resident;
(b) steps are taken to prevent resident entrapment, taking into consideration all potential zones of entrapment; and
(c) other safety issues related to the use of bed rails are addressed, including height and latch reliability. O. Reg. 79/10, s. 15 (1).

Order / Ordre :

The Licensee will ensure, where bed rails are used, the resident is assessed and his or her bed system evaluated in accordance with evidence-based practices to minimize risk to the resident and, all bed systems are evaluated according to the Health Canada Guidance Document for Adult Hospital Beds.

Grounds / Motifs :



**Ministry of Health and
Long-Term Care**

Order(s) of the Inspector

Pursuant to section 153 and/or
section 154 of the *Long-Term Care
Homes Act, 2007, S.O. 2007, c.8*

**Ministère de la Santé et
des Soins de longue durée**

Ordre(s) de l'inspecteur

Aux termes de l'article 153 et/ou
de l'article 154 de la *Loi de 2007 sur les foyers
de soins de longue durée, L.O. 2007, chap. 8*

1. The licensee has failed to ensure that where bed rails were used, the resident was assessed and his or her bed system evaluated in accordance with evidence-based practices, and if there were none, in accordance with prevailing practices to minimize risk to the resident; and to ensure that where bed rails were used, steps were taken to prevent resident entrapment, taking into consideration all potential zones of entrapment.

Review of documentation of the bed system entrapment assessments revealed assessments had not been completed for all bed systems in the home. There was no documented evidence that residents, who used bed rails, had been assessed.

Interview with the Environmental Service Supervisor (ESS) revealed that all the beds had not been assessed.

Interview with the Restorative Care Coordinator and the Administrator confirmed that the bed system entrapment assessments were not totally completed, that residents who used bed rails had not been assessed and no short term action was taken to prevent resident entrapment. [s. 15. (1)] (615)

This order must be complied with by /

Vous devez vous conformer à cet ordre d'ici le : Aug 26, 2015



**Ministry of Health and
Long-Term Care**

**Ministère de la Santé et
des Soins de longue durée**

Order(s) of the Inspector

Pursuant to section 153 and/or
section 154 of the *Long-Term Care
Homes Act, 2007*, S.O. 2007, c.8

Ordre(s) de l'inspecteur

Aux termes de l'article 153 et/ou
de l'article 154 de la *Loi de 2007 sur les foyers
de soins de longue durée*, L.O. 2007, chap. 8

REVIEW/APPEAL INFORMATION

TAKE NOTICE:

The Licensee has the right to request a review by the Director of this (these) Order(s) and to request that the Director stay this (these) Order(s) in accordance with section 163 of the Long-Term Care Homes Act, 2007.

The request for review by the Director must be made in writing and be served on the Director within 28 days from the day the order was served on the Licensee.

The written request for review must include,

- (a) the portions of the order in respect of which the review is requested;
- (b) any submissions that the Licensee wishes the Director to consider; and
- (c) an address for services for the Licensee.

The written request for review must be served personally, by registered mail or by fax upon:

Director
c/o Appeals Coordinator
Performance Improvement and Compliance Branch
Ministry of Health and Long-Term Care
1075 Bay Street, 11th Floor
TORONTO, ON
M5S-2B1
Fax: 416-327-7603



**Ministry of Health and
Long-Term Care**

Order(s) of the Inspector

Pursuant to section 153 and/or
section 154 of the *Long-Term Care
Homes Act, 2007, S.O. 2007, c.8*

**Ministère de la Santé et
des Soins de longue durée**

Ordre(s) de l'inspecteur

Aux termes de l'article 153 et/ou
de l'article 154 de la *Loi de 2007 sur les foyers
de soins de longue durée, L.O. 2007, chap. 8*

When service is made by registered mail, it is deemed to be made on the fifth day after the day of mailing and when service is made by fax, it is deemed to be made on the first business day after the day the fax is sent. If the Licensee is not served with written notice of the Director's decision within 28 days of receipt of the Licensee's request for review, this(these) Order(s) is(are) deemed to be confirmed by the Director and the Licensee is deemed to have been served with a copy of that decision on the expiry of the 28 day period.

The Licensee has the right to appeal the Director's decision on a request for review of an Inspector's Order(s) to the Health Services Appeal and Review Board (HSARB) in accordance with section 164 of the Long-Term Care Homes Act, 2007. The HSARB is an independent tribunal not connected with the Ministry. They are established by legislation to review matters concerning health care services. If the Licensee decides to request a hearing, the Licensee must, within 28 days of being served with the notice of the Director's decision, give a written notice of appeal to both:

Health Services Appeal and Review Board and the Director

Attention Registrar
151 Bloor Street West
9th Floor
Toronto, ON M5S 2T5

Director
c/o Appeals Coordinator
Performance Improvement and Compliance
Branch
Ministry of Health and Long-Term Care
1075 Bay Street, 11th Floor
TORONTO, ON
M5S-2B1
Fax: 416-327-7603

Upon receipt, the HSARB will acknowledge your notice of appeal and will provide instructions regarding the appeal process. The Licensee may learn more about the HSARB on the website www.hsarb.on.ca.



**Ministry of Health and
Long-Term Care**

Order(s) of the Inspector
Pursuant to section 153 and/or
section 154 of the *Long-Term Care
Homes Act, 2007, S.O. 2007, c.8*

**Ministère de la Santé et
des Soins de longue durée**

Ordre(s) de l'inspecteur
Aux termes de l'article 153 et/ou
de l'article 154 de la *Loi de 2007 sur les foyers
de soins de longue durée, L.O. 2007, chap. 8*

RENSEIGNEMENTS SUR LE RÉEXAMEN/L'APPEL

PRENDRE AVIS

En vertu de l'article 163 de la Loi de 2007 sur les foyers de soins de longue durée, le titulaire de permis peut demander au directeur de réexaminer l'ordre ou les ordres qu'il a donné et d'en suspendre l'exécution.

La demande de réexamen doit être présentée par écrit et est signifiée au directeur dans les 28 jours qui suivent la signification de l'ordre au titulaire de permis.

La demande de réexamen doit contenir ce qui suit :

- a) les parties de l'ordre qui font l'objet de la demande de réexamen;
- b) les observations que le titulaire de permis souhaite que le directeur examine;
- c) l'adresse du titulaire de permis aux fins de signification.

La demande écrite est signifiée en personne ou envoyée par courrier recommandé ou par télécopieur au:

Directeur
a/s Coordinateur des appels
Direction de l'amélioration de la performance et de la conformité
Ministère de la Santé et des Soins de longue durée
1075, rue Bay, 11^e étage
Ontario, ON
M5S-2B1
Fax: 416-327-7603

Les demandes envoyées par courrier recommandé sont réputées avoir été signifiées le cinquième jour suivant l'envoi et, en cas de transmission par télécopieur, la signification est réputée faite le jour ouvrable suivant l'envoi. Si le titulaire de permis ne reçoit pas d'avis écrit de la décision du directeur dans les 28 jours suivant la signification de la demande de réexamen, l'ordre ou les ordres sont réputés confirmés par le directeur. Dans ce cas, le titulaire de permis est réputé avoir reçu une copie de la décision avant l'expiration du délai de 28 jours.



**Ministry of Health and
Long-Term Care**

Order(s) of the Inspector

Pursuant to section 153 and/or
section 154 of the *Long-Term Care
Homes Act, 2007, S.O. 2007, c.8*

**Ministère de la Santé et
des Soins de longue durée**

Ordre(s) de l'inspecteur

Aux termes de l'article 153 et/ou
de l'article 154 de la *Loi de 2007 sur les foyers
de soins de longue durée, L.O. 2007, chap. 8*

En vertu de l'article 164 de la Loi de 2007 sur les foyers de soins de longue durée, le titulaire de permis a le droit d'interjeter appel, auprès de la Commission d'appel et de révision des services de santé, de la décision rendue par le directeur au sujet d'une demande de réexamen d'un ordre ou d'ordres donnés par un inspecteur. La Commission est un tribunal indépendant du ministère. Il a été établi en vertu de la loi et il a pour mandat de trancher des litiges concernant les services de santé. Le titulaire de permis qui décide de demander une audience doit, dans les 28 jours qui suivent celui où lui a été signifié l'avis de décision du directeur, faire parvenir un avis d'appel écrit aux deux endroits suivants :

À l'attention du registraire
Commission d'appel et de révision
des services de santé
151, rue Bloor Ouest, 9e étage
Toronto (Ontario) M5S 2T5

Directeur
a/s Coordinateur des appels
Direction de l'amélioration de la performance et de la
conformité
Ministère de la Santé et des Soins de longue durée
1075, rue Bay, 11e étage
Ontario, ON
M5S-2B1
Fax: 416-327-7603

La Commission accusera réception des avis d'appel et transmettra des instructions sur la façon de procéder pour interjeter appel. Les titulaires de permis peuvent se renseigner sur la Commission d'appel et de révision des services de santé en consultant son site Web, au www.hsarb.on.ca.

Issued on this 12th day of August, 2015

Signature of Inspector /

Signature de l'inspecteur :

Name of Inspector /

Nom de l'inspecteur : Helene Desabrais

Service Area Office /

Bureau régional de services : London Service Area Office