



**Ministry of Health and
Long-Term Care**

**Inspection Report under
the Long-Term Care
Homes Act, 2007**

**Ministère de la Santé et des
Soins de longue durée**

**Rapport d'inspection
prévue le Loi de 2007 les
foyers de soins de longue**

**Health System Accountability and Performance
Division**

Performance Improvement and Compliance Branch

**Division de la responsabilisation et de la
performance du système de santé**

**Direction de l'amélioration de la performance et de la
conformité**

Ottawa Service Area Office
347 Preston St, 4th Floor
OTTAWA, ON, K1S-3J4
Telephone: (613) 569-5602

Fax: (613) 569-9670

Bureau régional de services d'Ottawa
347, rue Preston, 4th étage
OTTAWA, ON, K1S-3J4
Téléphone: (613) 569-5602

Télécopieur: (613) 569-9670

Public Copy/Copie du public

Date(s) of inspection/Date(s) de l'inspection	Inspection No/ No de l'inspection	Type of Inspection/Genre d'inspection
Jul 21, 22, 25, 2011	2011_049143_0013	Critical Incident

Licensee/Titulaire de permis

REVERA LONG TERM CARE INC.
55 STANDISH COURT, 8TH FLOOR, MISSISSAUGA, ON, L5R-4B2

Long-Term Care Home/Foyer de soins de longue durée

HALLOWELL HOUSE
13628 LOYALIST PARKWAY, PICTON, ON, K0K-2T0

Name of Inspector(s)/Nom de l'inspecteur ou des inspecteurs

PAUL MILLER (143)

Inspection Summary/Résumé de l'inspection

The purpose of this inspection was to conduct a Critical Incident inspection.

During the course of the inspection, the inspector(s) spoke with The Director of Nursing and Assistant Director of Nursing.

During the course of the inspection, the inspector(s) Reviewed internal incident report, reviewed maintenance inspection reports and resident health care records.

The following Inspection Protocols were used in part or in whole during this inspection:

Accommodation Services - Maintenance

Findings of Non-Compliance were found during this inspection.

NON-COMPLIANCE / NON-RESPECT DES EXIGENCES

Definitions

WN – Written Notification
VPC – Voluntary Plan of Correction
DR – Director Referral
CO – Compliance Order
WAO – Work and Activity Order

Définitions

WN – Avis écrit
VPC – Plan de redressement volontaire
DR – Aiguillage au directeur
CO – Ordre de conformité
WAO – Ordres : travaux et activités



**Ministry of Health and
Long-Term Care**

**Inspection Report under
the Long-Term Care
Homes Act, 2007**

**Ministère de la Santé et des
Soins de longue durée**

**Rapport d'inspection
prévue le Loi de 2007 les
foyers de soins de longue**

Non-compliance with requirements under the Long-Term Care Homes Act, 2007 (LTCHA) was found. (A requirement under the LTCHA includes the requirements contained in the items listed in the definition of "requirement under this Act" in subsection 2(1) of the LTCHA.)

The following constitutes written notification of non-compliance under paragraph 1 of section 152 of the LTCHA.

Le non-respect des exigences de la Loi de 2007 sur les foyers de soins de longue durée (LFSLD) a été constaté. (Une exigence de la loi comprend les exigences qui font partie des éléments énumérés dans la définition de « exigence prévue par la présente loi », au paragraphe 2(1) de la LFSLD.

Ce qui suit constitue un avis écrit de non-respect aux termes du paragraphe 1 de l'article 152 de la LFSLD.

WN #1: The Licensee has failed to comply with O.Reg 79/10, s. 36. Every licensee of a long-term care home shall ensure that staff use safe transferring and positioning devices or techniques when assisting residents. O. Reg. 79/10, s. 36.

Findings/Faits sayants :

1. A resident was being transferred in a mechanical lift by Personal Support Workers. At the time of the transfer staff closed the legs on the left and proceeded to push the lift under the residents bed. During this, the lift got caught on a cord under the bed and tipped over. The resident landed on his bed and sustained no injuries. The lift was immediately taken out of service. An internal review reported that the staff did not follow the manufactures requirements for transferring the resident. The corporate health and safety staff member reported that the manufactures specifications indicate that the legs of the lift are to remain open during a resident transfer.

A service technician inspected all the lifts in the home and following the inspection all but one of the homes lifts were replaced.

Issued on this 25th day of July, 2011

Signature of Inspector(s)/Signature de l'inspecteur ou des inspecteurs